

CDC/ILB XPERT TB PROFICIENCY TESTING PROGRAM: INDIVIDUAL FINAL REPORT

2023-A

Country	Testing Site	PT-ID Number	Page Number
El Salvador	Hospital Nacional Rosales	22003	1/3

Instrument SN	Last Calibration Date	Result Submission Date
708249	12-Dec-2022	06-Jun-2023
803507	12-Dec-2022	06-Jun-2023

Site Results Testing with MTB Ultra


	MTB Detected	RIF Resistance	Score
Sample ID: 2023-A-1			
Expected Results-MTB/RIF	Medium	Not Detected	
Expected Results-Ultra	Medium	Not Detected	
All Participants' Consensus Results-MTB/RIF	Medium	Not Detected	
All Participants' Consensus Results-Ultra	Medium	Not Detected	
Hospital Nacional Rosales	Low	Not Detected	20
Sample ID: 2023-A-2			
Expected Results-MTB/RIF	Medium	Detected	
Expected Results-Ultra	Medium	Detected	
All Participants' Consensus Results-MTB/RIF	Medium	Detected	
All Participants' Consensus Results-Ultra	Medium	Detected	
Hospital Nacional Rosales	Low	Detected	20
Sample ID: 2023-A-3			
Expected Results-MTB/RIF	Not Detected	N/A	
Expected Results-Ultra	Not Detected	N/A	
All Participants' Consensus Results-MTB/RIF	Not Detected	N/A	
All Participants' Consensus Results-Ultra	Not Detected	N/A	
Hospital Nacional Rosales	Not Detected	N/A	20
Sample ID: 2023-A-4			
Expected Results-MTB/RIF	Medium	Not Detected	
Expected Results-Ultra	Medium	Not Detected	
All Participants' Consensus Results-MTB/RIF	Medium	Not Detected	
All Participants' Consensus Results-Ultra	Medium	Not Detected	
Hospital Nacional Rosales	Medium	Not Detected	20
Sample ID: 2023-A-5			
Expected Results-MTB/RIF	Medium	Detected	
Expected Results-Ultra	Medium	Detected	
All Participants' Consensus Results-MTB/RIF	Medium	Detected	
All Participants' Consensus Results-Ultra	Medium	Detected	
Hospital Nacional Rosales	Medium	Detected	20
	Percentage	Satisfactory/Unsatisfactory	
FINAL SCORE	100	Satisfactory	

CDC/ILB XPERT TB PROFICIENCY TESTING PROGRAM: INDIVIDUAL FINAL REPORT

2023-A

Country	Testing Site	PT-ID Number	Page Number
El Salvador	Hospital Nacional Rosales	22003	2/3

Comments



CDC/ILB XPERT TB PROFICIENCY TESTING PROGRAM: INDIVIDUAL FINAL REPORT

2023-A

Country	Testing Site	PT-ID Number	Page Number
El Salvador	Hospital Nacional Rosales	22003	3/3

Report Issued by: Patricia Hall-Eidson

Date: 21-Jul-2023

Documentation of Report Review

We the undersigned, have read and reviewed the above Xpert TB Proficiency Test Individual Performance Report. Expected results for which testing site's submitted results are scored against are determined by the US Centers for Disease Control and Prevention (CDC)/ International Laboratory Branch (ILB) sample validation testing procedures prior to panel shipment. If the final score was less than 100%, we completed a root cause analysis. * The Individual Final Report and any attachments must be filed and retained as documentation. If needed, corrective action documentation should also be completed and filed alongside the Individual Performance Report for reference.

Our identifiable individual site results are confidential and are only shared with our Proficiency Test Country Coordinators (PTCC) at the National or Regional level, the testing site, or other PTCC-designated stakeholders. Each PTCC serves as a subcontractor to the Xpert TB Proficiency Test Program for the following duties: site enrollment, panel receipt, panel distribution to individual testing sites, result submission assistance, report distribution assistance and corrective action follow-up in accordance with the country's National TB Program guidelines. Unidentified site results may be shared for XTPT program review and improvement purposes upon request to, and with approval from, CDC/ ILB. Questions or comments concerning this report may be submitted to the PTCC or to the CDC/ ILB using the Contact Information listed below.

Testing Personnel _____ Date: _____

_____ Date: _____

Laboratory Manager (or Designee) _____ Date: _____

* If corrective action forms are not available from your country's National TB or Quality Assurance Program, corrective actions forms may be downloaded from ePT (<http://ept.systemone.id/>).

Proficiency Test Country Coordinators (PTCC)

Marlin Renderos mrenderos@comisca.net	Yanira Meléndez yani_melendez07@hotmail.com
---	---

Contact Information

Xpert TB Proficiency Test Program
xtpt@cdc.gov

Kyle DeGruy

Program Coordinator, TB and Clinical Monitoring Team
International Laboratory Branch
Division of Global HIV and TB
Center for Global Health
US Centers for Disease Control and Prevention
gsz4@cdc.gov

Patricia Hall-Eidson

Team Lead, TB and Clinical Monitoring Team
International Laboratory Branch
Division of Global HIV and TB
Center for Global Health
US Centers for Disease Control and Prevention
igg5@cdc.gov